

VERIFICATION OF CALIBRATION REPORT*of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program***MAY 30 2009**

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130116</u> ✓	
A	Name <u>Kenneth Parker</u>	ID# <u>5578</u>	Date <u>5/28/09</u> ✓
	Agency <u>McGrath Public Safety</u>	Phone # <u>524-3075</u>	
	Instrument Location <u>AST McGrath</u> ✓		
B	Alco S/N <u>77111</u>	Target Value <u>.078</u>	High Pressure <u>700</u>
	Alco Test Values	<u>.075</u> ✓ 1 st Alco	<u>.075</u> ✓ 2 nd Alco
	Signature <u>[Signature]</u> ✓		
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 4th day of June 2009.

[Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130116

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

MAY 28, 2009

OPERATOR'S NAME:

PARKER/KENNETH W

OPERATOR'S NUMBER: 5578

SUBJECT'S LAST NAME:

PARKER

SUBJECT'S FIRST NAME/MI :

KENNETH W

O.L. #: 130116

DEPT/AGENCY: MCS1

CASE/REPORT: 1

TEST TYPE: T

ALCO TARGET VALUE: 0.08

ALCO S/N: 77111

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.77 in

ALCO TARGET .077 10:43

BLANK TEST .000 10:44

INTERNAL STANDARD VERIFIED 10:44

ALCO TV 29.77 in .075 10:45

BLANK TEST .000 10:45

SUBJECT SAMPLE .000 10:46

BLANK TEST .000 10:47

ALCO TV 29.77 in .075 10:47

BLANK TEST .000 10:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

MAY 28, 2009

TIME 10:39

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER:

BREATH TUBE:

BAROMETER: 29.77 in

FLOW DETECTOR: OKAY

PUMP

WASH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~